Maximum Safety

Rotary Endodontic Files

Clinical Card

Maximum Safety

1. **Choose the first rotary file** according to the largest k-file to bind at primary working length.
2. **Pre-curving the file** assures conservative access cavity preparation and easier access in challenging cases.
3. **Irrigate thoroughly** with Sodium Hypochlorite before and between files, to ensure high quality disinfection and prevent file breakage.
4. **Minimal size to prepare a narrow canal**: ISO #30, to assure sufficient irrigation at WL.

Difficult
Initial size: ≤#10
Prepare to #30

Medium
Initial size: #15
Prepare to #30

Simple canals
Initial size: #20-40
Increase by 3 sizes (#30+)

Retreatment
Remove main bulk,
Prepare to #30-40,
as needed



MS Challenge #10/04	MS Trail #15/02-04	MS 20 #20/04	MS 25 #25/04	MS 30 #30/04	MS 35 #35/04-03	MS 40 #40/02	MS Retreatment #25/06-04-02
300-400 RPM	300-400 RPM	300-400 RPM	300-400 RPM	300-400 RPM	300-400 RPM	300-400 RPM	600-1000 RPM
1.5 N.cm	1.5 N.cm	2.5-3 N.cm	2.5-3 N.cm	2.5-3 N.cm	2.5-3 N.cm	2.5-3 N.cm	3 N.cm

osada ENDO

Orders: 03-7979620 | Customer service: 03-7979444 | www.ms-endo.com